



EMSL Analytical, Inc.

464 McCormick Street San Leandro, CA 94577
Tel/Fax: (510) 895-3675 / (510) 895-3680
<http://www.EMSL.com> / sanleandrolab@emsl.com

EMSL Order: 091706444
Customer ID: SMGS78C
Customer PO:
Project ID: Oroville EM OP

Attention: Ursula Parker
Butte County AQMD
629 Entier Avenue, Suite 15
Chico, CA 95928

Phone: (530) 332-9400
Fax: (530) 332-9414
Received Date: 4/ 5/2017 9:30 AM
Analysis Date: 04/06/2017
Collected Date: 04/04/2017

Project: BCAQMD 4/4/17 (Oroville EM OP)

Test Report: Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM) Performed by CARB-Modified AHERA

Sample	Location	Volume (Liters)	Area Analyzed (mm ²)	Non Asb	Asbestos Types	#Structures		Analytical Sensitivity (S/cc)	Asbestos Concentration	
						≥0.5μ < 5	≥5μ		(S/mm ²)	(S/cc)
BCAS08 091706444-0001	TEM SAMPLE (CARB)	2388.30	0.1690	23	None Detected			0.0010	<5.90	<0.0010
Non-asbestos structures reported contains a contribution from non-regulated amphiboles.										
F-1 091706444-0002	FIELD BLANK 1	0.00	0.1300	0	None Detected			N/A	<7.70	N/A
F-2 091706444-0003	FIELD BLANK 2	0.00	0.1300	0	None Detected			N/A	<7.70	N/A

Analyst(s)
Matthew Batongbacal (3)


Matthew Batongbacal
or other approved signatory

EMSL maintains liability limited to cost of analysis. This report relates only to the samples reported above and may not be reproduced, except in full, without written approval by EMSL. EMSL bears no responsibility for sample collection activities or analytical method limitations. Interpretation and use of test results are the responsibility of the client. This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government. Results reported in both structures/cm3 and structures/mm2 are dependent on the volume of Airsampled and measured by non-laboratory personnel are not the responsibility of EMSL and are not covered by the laboratory's NVLAP accreditation. Samples received in good condition unless otherwise noted. Estimated accuracy, precision and uncertainty data available upon request.

Samples analyzed by EMSL Analytical, Inc San Leandro, CA NVLAP Lab Code 101048-3, WA C884

Initial Report From: 04/07/2017 07:12:57